

VOP-026 Šternberk, s.p.
division VTÚPV Vyškov

V. Nejedlého 691, 682 03 Vyškov, Czech Republic
Product Certification Body

CERTIFICATE
N° VTÚPV – 072 / 2009

Applicant: **Xuzhou Jingying Electronics Technology Co., Ltd.**
No.1 Huangshan Rd. Tongshan Economic Development Zone Xuzhou
Jiangsu, China

Product: **Temperature Data Logger**

Model: **RC**

Manufacturer: **Xuzhou Jingying Electronics Technology Co., Ltd.**
No.1 Huangshan Rd. Tongshan Economic Development Zone Xuzhou
Jiangsu, China

Rating and principal characteristics: Power Supply: One lithium battery; Accuracy: ± 1 °C;
Temperature Measuring Range: -40 to +70 °C.

Test results are described in the Test Reports No.:
SCC(09)-113-10-EMC (tests made by CHINA CEPREI (SICHUAN) LAB.).

The sample of tested product conforms with the requirements of the following harmonized standards:

- EN 55022: 2006;
- EN 55024: 1998 + A1: 2001 + A2: 2003;
- EN 61000-3-2: 2006;
- EN 61000-3-3: 1995 + A1: 2001 + A2: 2005.


This certificate is valid until: **10.07.2014**

After preparation of the necessary technical documentation as well as the conformity declaration, the required CE marking can be affixed on the product. Other relevant directives have to be observed.
The CE mark shall appear according to this sample:



Vyškov 10.07.2009

Tel./Fax: +420 517 303 603
e-mail: m.bezdek@vtupv.cz
<http://www.vop.cz>


Milan Bezdek
Certification Head

